

Let's Get Started...

User Guide

With condition guides written by chartered
physiotherapist and women's health specialist,
Amanda Savage MCSP MSt




Welcome!

Thank you for choosing Kegel8® Ultra 20. This personal therapeutic device is designed to help you strengthen your pelvic floor in the comfort of your own home. Created especially for women, the Kegel8® Ultra 20 is a safe and effective neuromuscular stimulator (STIM) that works by producing electrical pulses that gently stimulate the muscles and nerve fibres in your pelvic area.

What Are the Benefits of Having a Stronger Pelvic Floor?

The pelvic floor is a sling of muscles that run from your pelvic bone (in the front) to your tailbone (in the back). Think of these muscles like a hammock that supports your reproductive organs, bladder and lower bowels.

Various life circumstances can weaken these muscles, such as pregnancy, childbirth, excessive weight gain, surgery, or ageing. By using the



Kegel8® Ultra 20, you will be effortlessly guided through a series of STIM-induced pelvic floor (kegel) exercises that will assist you in gradually restrengthening your pelvic floor.

Advice For First Time Users

If this is your first time using a STIM device, you may be a bit apprehensive; that's totally normal. The Kegel8® Ultra 20 has been designed with the utmost attention to quality and comfort. Simply read this handy guide completely and follow the instructions for proper use, and you will have a positive experience.

If at any moment during your use of the Kegel8® Ultra 20 you have a question or concern, we are happy to assist you. You can reach one of our advisors by visiting our website www.kegel8.com/contactus.

Welcome!	4
What Are the Benefits of Having a Stronger Pelvic Floor?	4
Advice For First Time Users	5
Who Can Use This Device?	8
Who Should Not Use This Device	9
Warnings and Safety Precautions	10
General Warnings	10
Hygiene	11
Quick Start Guide	12
Kegel8® Ultra 20 Checklist	12
Setting Up and Using Your Kegel8® Ultra 20	13
Using the Electrode Pads	16
List of Pre-Set Programmes	17
Programme Descriptions	19
Sensitivity Testing	19
Pelvic Floor Exercises	20
Intimate Sensation	21
Stress Incontinence, Overactive Bladder, Nocturia and Frequency	22
Postnatal Recovery and Post-Surgical Recovery	24

Bowel Incontinence	25
Pain Relief	26
How to Create a Custom Programme	27

Tailored Condition Guides by Physiotherapist Amanda Savage

1. More Effective Pelvic Floor Exercises	30
2. Bladder Weakness & Stress Incontinence	39
3. Uncontrollable Gas/Wind & Bowel Control	47
4. Menopause, Perimenopause & Postmenopause	54
5. Pelvic Organ Prolapse	63
6. Pelvic Pain	71
7. Postpartum Recovery	81
8. Vaginal Looseness & Lack of Sensation	88
9. Rehabilitation After Gynaecological Surgery	96

Frequently Asked Questions	104
---	------------

Product Support	112
------------------------------	------------

Who Can Use This Device?

The Kegel8® Ultra 20 has been designed to help women who wish to address the following issues:

Stress Urinary Incontinence – A condition that causes urine leakage when pressure is applied to the bladder.

Overactive Bladder – A frequent and sudden urge to urinate that is difficult to control.

Faecal Incontinence – A condition where bowel movements, wind and gas cannot be controlled, leading to stool leakage.

Pelvic Pain and Back Pain – Because of the close relationship between the lower back and the pelvic region, strengthening the pelvic floor can relieve this type of discomfort.

Natural Ageing – Age-related loosening and weakening of the vaginal and pelvic muscles.

Painful Intercourse – Common causes of painful intercourse are weak vaginal walls, overactive vaginal walls, and chronic dryness, which can be alleviated by STIM-supported kegel exercises.

Lack of Sensation – Women with a weak pelvic floor may experience less sexual sensation.

Who Should Not Use This Device



The Kegel8® Ultra 20 should not be used under the following circumstances in a home setting without the guidance of a medical professional:

- If a woman is pregnant.
- If she's fitted with a demand-style cardiac pacemaker.
- If an individual has an undiagnosed pain condition.
- If a person has an undiagnosed skin, vaginal, or rectal condition.
- If a woman has an active pelvic cancer.
- If a person fails to understand the user instructions that accompany the Kegel8® Ultra 20.

The Kegel8® Ultra 20 should never be used under the following circumstances:

- If a person is immersed in or near water.
- If a person is connected to a high-frequency device.
- If a person is operating a motor vehicle or potentially dangerous machinery.

Warnings and Safety Precautions

Before you begin using the Kegel8® Ultra 20, make sure to completely read the following information to ensure safe and proper use of your device.

General Warnings

- Use the Kegel8® Ultra 20 only as directed.
- Keep the device out of reach of children.
- Do not immerse this device in water or any other liquid or expose it to rain or moisture.
- Do not use the Kegel8® Ultra 20 on any part of your body above your pelvis without guidance from a qualified physician.
- If you are uncertain about how to properly use this device, contact one of our dedicated advisors at the following link: www.kegel8.co.uk/contactus.
- Only use a 9v battery to power your Kegel8® Ultra 20. Avoid using Ni-cad rechargeable batteries.
- Although this device can be used independently, it's advised to use the Kegel8® Ultra 20 under the supervision of a physiotherapist or continence advisor if possible.

Hygiene


- Never share probes or electrodes with other people. These items are for personal use by a single person only.
- Always clean probes thoroughly after every use and store them away dry.
- Before applying electrode pads, make sure skin is clean and free from any oils, lotions or moisturisers.

Cleaning and Storage

After use, clean your probe with warm water, optionally spray on an antibacterial spray, and then wipe dry. With the lead wires still attached, store your probe with the Kegel8® Ultra 20 in the convenient carrying case which came with your unit.

Quick Start Guide

This section of your user manual is meant to be a go-to reference to help you with setting up your Kegel8® Ultra 20 device and getting comfortable with using it. We suggest reading through the entire manual at least once before moving forward with the following instructions.




Kegel8® Ultra 20 Checklist

Your package should contain the following items:

- 1 x Kegel8® Ultra 20 device (A)
- 1 x Kegel8® Ultra 20 Probe (B)
- 1 x 9 Volt Battery (C)
- 2 x Kegel8® Ultra 20 Lead Wires (D)
- 4 x Kegel8® Ultra 20 Electrode Pads (E)
- 1 x Kegel8® Ultra 20 User Manual


Setting Up and Using Your Kegel8® Ultra 20




1. Insert the 9 Volt Battery - Remove the back cover of the Kegel8® Ultra 20. Take one 9-volt battery (one has been included with your device.) Insert the battery into the device, and replace the back cover.



2. Prepare Your Probe/Electrodes - Wash your probe in warm, soapy water (body safe/ intimate soap) or use a gentle antibacterial cleanser. Then rinse and dry thoroughly. This does not need to be done with your electrodes, but ensure the skin is clean and dry instead.




3. Connect the Probe/Electrodes to the Kegel8® Ultra 20 Device - Insert the lead wire into Channel A at the top left of the device and connect your probe/electrodes to the lead wire connectors. Turn on your Kegel8® Ultra 20 - Press the on/off button once to switch on.



4. Insert Your Probe/Attach Your Electrodes - Apply a small amount of Kegel8 Conductivity Gel, on the metal parts of the probe. Insert the probe into your vagina, with the metal sides facing “hip-to-hip”. Avoid placing too deep; you should be able to touch the base of the probe with your finger tip, approx. one knuckles depth inside the vagina entrance. If using electrodes, place in the desired position and ensure they are stuck effectively.



5. Select a Programme - Choose an exercise programme by pressing the PRG button until the screen shows the programme you wish to use.



6. To proceed with a selected programme, press the A+ button to increase the current intensity (displayed as mA = milliamps). If the sensation is too strong, you can

decrease it by pushing the A- button. The amount of current needed to stimulate pelvic floor muscles varies from person to person. Starting from 0 mAs, you should increase the mA power to a level where you can feel the effects of the electrical pulse without it causing any discomfort.

TIP: To pause a programme, press the P button. When you're ready to continue press the + button.

For a sensation test or a neuromodulation programme, start at 0 and increase the mA level until you feel a slight tingling/knocking sensation. This feeling indicates that the programme is starting to work. You should then keep the device at this level for your sensation test and the first week of your routine, before increasing the mA level by 1-2 mA's every 1-2 weeks. Setting the mA level too high, is unnecessary and will cause the pelvic floor or surrounding muscles to over work and may cause discomfort.

For a muscle exercise programme you can gradually increase until you feel the muscles contract and relax with the electrical cycles. The sensation should never be painful. Some programmes have different phases, if you feel a decrease in intensity as the phase changes, just use the + button to increase the mAs back to a comfortable level.

TIP: For first time users, start with P01 - Sensation Test 1. This will ensure that you are able to feel the current safely. If you cannot feel the current, perform P01 at 50mA once a day for 4 weeks before progressing onto P02.

Using the Electrode Pads


The electrode pads that came with your device can be used as an alternative to the internal probe, below are options for where you can place the electrode pads. **Please note, ankle positioning should only be used for P09 (TTNS), it is not appropriate for any other programme.**



When using 2 pads, simply connect a lead wire to a pair of electrode pads and insert the other end of the lead wire into Channel A on the top left of your device. Place on the skin and use Channel A controls to change the intensity of the current. you will not need to turn it up very far to feel a sensation.

When using 4 pads, use 2 lead wires, with one in Channel A and one in Channel B. Use the corresponding controls to turn up the intensity of the current through each pair of pads individually; they do not have to be the same intensity.

Where to Apply Electrode Pads



Position 1
Abdomen






Position 2
Sacrum



Position 3
Ankle

List of 20 Pre-Set Programmes

-  Green: Sensation test - for testing your sensation levels
-  Lilac: Muscle exercises - for strengthening your pelvic floor muscles
-  Pink: Neuromodulation programmes - for calming overactive nerves, circulation, relaxation and pain relief

All programmes can be performed with the internal probe. Alternatively, the electrode pads can be used in the positions detailed below.

Prog. No.	Screen Code	Programme Name	Electrode Positioning	Time (mins)
P01	RCV1	Sensation Test 1	1,2 or 3	30
P02	RCV2	Sensation Test 2	1,2 or 3	60
P03	EXE1	Pelvic Floor Muscle Exercises 1	2	45
P04	EXE2	Pelvic Floor Muscle Exercises 2	2	30
P05	EXE3	Pelvic Floor Muscle Exercises 3 (maintenance programme)	2	45
P06	SIMP	Sensation Improvement	2	20

Continued overleaf...

P07	ENDU	Pelvic Floor Muscle Exercises 4 - Endurance	2	60
P08	SENS	Sensory modulation (TENS)	2	60
P09	OAB	Overactive Bladder neuro-modulation	2 or 3	20
P10	STR1	Stress incontinence 1	2	40
P11	STR2	Stress Incontinence 2	2	40
P12	MIXDS	Mixed Incontinence	2	30
P13	FTRN	Functional Continence Training (with skin electrodes)	2	9
P14	PNC1	Postnatal / Post-Surgical Care 1	2	35
P15	PNC2	Postnatal / Post-Surgical Care 2	2	30
P16	BOW1	Bowel (Faecal) Incontinence 1	2	10
P17	BOW2	Bowel (Faecal) Incontinence 2	2	40
P18	PAIN	Chronic Pain Relief (with skin electrodes)	1 or 2	∞
P19	CIRC	Pelvic Floor Circulation	1 or 2	60
P20	RELX	Pelvic Floor Relaxation	2	20

Programme Descriptions

Sensitivity Testing

P01 - Sensation Test 1 Use this programme to test that you can feel the current safely.

Use the + button on the same side as you are plugged in to the machine to turn up the current (measured in mA). If you can feel a sensation (usually a warm tickling) from the internal probe below 50mA progress to the second sensation test P02.

If you cannot feel a tickling sensation below 50mA it can be a sign of nerve damage or difficulty conducting messages. This is quite common after childbirth or surgery.

Use P01 as your treatment programme, to help the nerves recover, at 50mA (or less), once a day for 4 weeks before progressing onto P02. If you cannot feel sensation returning after 1 month of use please consult a medical professional for advice.

P02 - Sensation Test 2 This second test introduces new frequencies. If you can feel a sensation from the internal probe below 50mA, progress to one of the Kegel8 treatment programmes (P03-P20).

If you cannot feel a sensation below 50mA it can be

a sign of nerve damage. Use P02 at 50mA once a day for a further 4 weeks.

If you do not feel sensation returning contact your doctor or physiotherapist for advice.

Pelvic Floor Exercises

P03 - Pelvic Floor Muscle Exercises 1 is a 45-minute muscle exercise programme. Use to improve your awareness of how your muscles contract and relax. You will feel a short (6 second) muscle contraction then a long recovery rest before the next muscle contraction.

P04 - Pelvic Floor Muscle Exercises 2 is a 30-minute muscle exercise programme similar to P03. Use to improve your muscle endurance with a longer contraction time (8 seconds).

P04 changes through 4 phases, moving between lower and medium frequencies.

P05 - Pelvic Floor Muscle Exercises 3 is a 45-minute muscle exercise programme ideal for a weekly maintenance session. The work/rest cycles are shorter so you will do many contraction cycles. 2 phases of lower and medium frequencies.

Intimate Sensation

P06 - Sensation Improvement is a shorter 20-minute muscle exercise session of high frequency electrical current. Use to improve sensory awareness and create strong pelvic floor muscle contractions.

P07 - Pelvic Floor Muscle Exercises 4 is a long, one-hour, muscle exercise programme. Use to improve endurance once muscles are stronger and only if they do not fatigue easily. You will feel repeated short muscle contraction cycles (5 seconds contract, 5 seconds relax).

P08 - Sensory Modulation (TENS) is a one-hour nerve-calming neuromodulation programme to calm oversensitive nerves. There are 2 phases. The first 45-minute phase of low frequency continuous current, also known as TENS, will improve blood flow to the pelvic floor muscles and has a soothing effect on nerves. You can apply this current using an internal probe or with electrode pads on the lower back/sacrum area. You do not need to feel the muscle contracting, aim to relax completely. If using the internal probe, you can complete the full programme by continuing on to phase 2, for 15 minutes of high frequency stimulation in a work/rest pattern to exercise your pelvic floor muscles.

Stress Incontinence, Overactive Bladder, Nocturia and Frequency

P09 - Overactive Bladder Neuromodulation is a 20-minute nerve-calming neuromodulation programme to calm oversensitive nerves. A low frequency current has a soothing effect on the bladder nerves. Use this for overactive bladder condition or symptoms of urgency, frequency and urge incontinence.

It can be applied close to the pelvic floor with the probe. It can also be applied with skin electrodes on the lower back/sacrum area, or the ankle for TTNS (details on page 47). See page 16 for pad placements.

You will feel the current in cycles of 10 seconds with 5 second rests. You do not need to feel the muscles contracting. Use once a day for the first 6 weeks, after which you can increase to twice a day if comfortable.

P10 - Stress Incontinence 1 is a 40-minute muscle exercise programme.

Phase 1: 10 minutes of continuous low frequency current improves awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes medium frequency current to stimulate the fast and slow pelvic floor muscle fibres

by contracting for 5 seconds then relaxing for 5 seconds in cycles.

P11 - Stress Incontinence 2 is a 40-minute muscle exercise programme similar to P10. Both phases actively exercise the muscles by contracting for 5 seconds then relaxing for 5 seconds in cycles.

P12 - Mixed Incontinence is a 30-minute programme in two even phases of stimulation in a work/rest pattern (5 seconds work, 8 seconds rest). Phase 1 is low frequency current which helps calm pelvic and bladder nerves. Phase 2 is high frequency stimulation which promotes development of the pelvic floor muscles. This can be applied with 4 skin electrode pads on the lower back/sacrum area.

P13 - Functional Continence Training is a muscle exercise programme for application with 4 skin electrode pads. It requires you to stand, sit and walk while running the programme. See the Technical Instructions for Use (IFUS) for indepth instructions for this programme.

Postnatal Recovery and Post-Surgical Recovery

P14 - Postnatal / Post Surgical Care 1 is a 35-minute

gentle start pelvic floor muscle exercise programme for muscles that are vulnerable, for example after a birth, gynaecological surgery or from a prolapse.

A vaginal probe can be used 12 weeks after birth or surgery.

Phase 1: 20 minutes of continuous low frequency current improves awareness of the muscle position, blood flow and nerve sensitivity. Relax in this phase.

Phase 2: 15 minutes high frequency current to stimulate the pelvic floor muscle fibres by contracting for 6 seconds then relaxing for 18 seconds in cycles (1:3). This pattern prevents fatigue.

P15 - Postnatal / Post Surgical Care 2 is also a gentle start pelvic floor muscle exercise programme for muscles that are vulnerable, for example after a birth, gynaecological surgery or when coping with the pressure of prolapse.

It is very similar to P14. In the second phase slightly longer muscle contraction time (8 seconds) with a similar 1:3 rest cycle to reduce muscle fatigue.

Bowel Incontinence

P16 - Bowel (Faecal) Incontinence 1 is a 10-minute muscle exercise programme.

If you have weak posterior muscles or anal sphincter this programme can be applied using a vaginal probe or anal probe.

You will feel repeated short muscle contraction cycles (5 seconds contract, 5 seconds relax) in 2 phases of different frequency current

P17 - Bowel (Faecal) Incontinence 2 is a 40-minute muscle exercise programme.

If you have weak posterior muscles or anal sphincter this programme can be applied using a vaginal probe or anal probe, or with 4 skin electrode pads on the lower back/sacrum area.

This programme is similar to P16 with current at a higher frequency. Similar short contraction cycles (5 seconds contract, 5 seconds relax) in one phase of 40 minutes.

Pain Relief

P18 - Chronic Pain Relief is a 10-minute nerve-calming neuromodulation programme to calm oversensitive nerves.

This programme is a high frequency current (known as high frequency TENS) in a cycle pattern (5 seconds current flow, 15 seconds rest), for

use with one or 2 pairs of skin electrode pads. This programme will be helpful for chronic pain problems such as back pain, period pain, pelvic pain and vulvodynia.

P19 - Pelvic Floor Circulation is a 60-minute nerve-calming neuromodulation programme to calm oversensitive nerves. This programme is a low frequency continuous current (known as low frequency acupuncture-like TENS) in one phase of 60 minutes.

This programme will be helpful for muscle cramps or tight pelvic muscles. Use this programme with a vaginal probe or skin electrode pads.

P20 - Pelvic Floor Relaxation is a 20-minute nerve-calming neuromodulation programme for oversensitive nerves using an internal probe. This short programme has 2 phases of low frequency continuous current.

Phase 1: 10 minutes of current in cycles (3 seconds current flow, 7 seconds no current flow). You can concentrate on fully relaxing any tension in your pelvic floor muscles in cycle with the current.

Phase 2: 10 minutes continuous low frequency current. Aim to completely relax your body and pelvic floor muscles.

How to Create a Custom Programme

Many women find that the pre-set programmes built into the Kegel8® Ultra 20 are ideal for treating their pelvic floor problems; however, in some situations, a medical professional may advise using a different type of programme. The Kegel8 Ultra 20 has a custom programme feature that allows doctors/clinicians to create up to 3 programmes designed specifically for you. For details about how a custom programme is created, refer to the Technical Instructions for Use (IFUS) that came with your device.

Tailored Condition Guides

Address Your Specific Issue(s)



The following 9 condition guides were written by Amanda Savage, a specialist physiotherapist.

Amanda is a member of the Professional Network of Pelvic Obstetric & Gynaecological Physiotherapy and a graduate of the University of Cambridge.

These guides combine several of the pre-set Kegel8® Ultra 20 programmes into structured routines for addressing specific problems. You can think of them as targeted therapy once you're familiar with using your device.

Use the following list to easily navigate to the condition guide that best fits your issue.

Condition Guides

- 1 More Effective Pelvic Floor Exercises 30
- 2 Bladder Weakness & Stress Incontinence 39
- 3 Uncontrollable Gas/Wind & Bowel Control 47
- 4 Menopause, Perimenopause & Postmenopause 54
- 5 Pelvic Organ Prolapse 63
- 6 Pelvic Pain 71
- 7 Postpartum Recovery 81
- 8 Vaginal Looseness & Lack of Sensation 88
- 9 Rehabilitation After Gynaecological Surgery 96



1. More Effective Pelvic Floor Exercises

Many women practise pelvic floor kegel exercises to improve bladder or bowel control and pelvic support. If they've found these exercises difficult to perform, the use of a STIM machine like the Kegel8® Ultra 20 helps with attaining results easier and faster.

STIM machines are ideal for people who really don't know where their muscles are or how to activate them. They're also good for very weak muscles that can only contract a few times before they fatigue. STIM devices are also useful for people who find that they get easily distracted when doing traditional kegel exercises; using a stimulation device makes it easier to concentrate.

These are some of the primary benefits of using a STIM device to assist with kegel exercises:

- Sensations created by the electrical impulses help your brain to correctly identify the location of your pelvic floor muscles.
- A STIM device will stimulate the muscles to work (contract). It's possible to change the settings to favour different types of muscle fibre, and to do exercises in different ways, for example, longer or shorter muscle holds.
- You can practise contracting your muscles along with the machine to prepare for “real life” scenarios where you won't have access to the device.
- The machine can help you to “hold” a muscle contraction while you concentrate on learning to breathe at the same time. You can practise using pre-set programmes or set up a custom programme of your own.
- You can use the skills you learn when working with the machine to improve your pelvic floor practise when exercising without the machine.



Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If you're worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist.

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P03

P03 features long recovery rests between muscle contractions.

Phase 1: 15 minutes low frequency.

Phase 2: 30 minutes higher frequency.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Only use alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P03

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or experiencing any type of irritation.
- If your internal tissues are tender or easily

irritated, continue an alternate day pattern.

1 2 3 4 Week 3 - 4: Use P04

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.

Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation. Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Notice your breathing - make sure that you ARE breathing even when you're contracting your muscles.

Second Month 1 2 3 4

Goal: Improve your muscle (and brain!) skills

1 2 3 4 Week 1 - 4: Use P10

Phase 1: First enjoy 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the entire 5 seconds. Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Change the positions in which you practise. You could do one session lying, one session sitting, and one session standing (but stand STILL - don't try walking about!)
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive).

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases (these are fairly short cycles now, 5 seconds contract, 5 seconds rest).

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

TIPS

- Aim to hold your muscles with your mind as well as with the machine for the entire contraction part of the cycle.
- Make sure you are not fading before the contraction is over.
- Remember to exercise in different positions (lying, sitting and standing). Especially standing now as this is where life takes place!

Or: work with P04 again

This month working with P04 try some more advanced skills by changing your focus throughout the session:

- Try to hold 2 seconds longer than the machine contraction.
- Be extra aware of completely relaxing in the rest part of the cycle.
- Are you breathing all through the cycle?
- During the contractions - could you move your arms (try some biceps curls, or backstroke) and hold your muscles at the same time and breathe?!
- Remember to exercise in different positions (lying, sitting and standing). Especially standing now as this is where life takes place!

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the Maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Phase 1: 15 minutes.

Phase 2: 30 minutes.

Remember to join in with some of the contractions to keep your brain well connected to the pelvic floor muscle work!



2. Bladder Weakness and Stress Incontinence

Stress incontinence is leakage that occurs when the bladder is under pressure or high impact, such as when you cough, sneeze, laugh or jump. Exercising with a STIM device makes the pelvic floor muscles contract and release to encourage them to grow and have a better blood and nerve supply. Follow our guided 12-week programme using the internal probe to directly activate the pelvic floor muscles or if you prefer, indirectly using skin electrodes (place on the sacrum, see placement on page 16).

Urge urinary incontinence is the urgent and frequent need to urinate or experience leakage before you can make it to the toilet. You will also benefit from our 12-week pathway of stimulation to strengthen your pelvic floor muscles.

If your bladder feels predominantly “jumpy” or “irritable”, you can also use your stimulation machine to “calm” the nerves to the bladder. The pathway will guide you to introduce this at week 4 once you have established your muscle training routine. Or, if this is your primary concern, you can skip to page 47 for guidance on how to apply a neuromodulation programme with skin electrodes at the ankle (TTNS).

Benefits of Using a Stimulation Machine

- Sensations created by the electrical impulses help your brain to correctly identify the location of your pelvic floor muscles.
- You can practise contracting your muscles along with the machine to prepare for “real life” scenarios where you can’t access your device.
- The machine can help you to “hold” a muscle contraction while you concentrate on learning to breathe and move at the same time.
- You can use the skills you learn when working with the machine to improve your pelvic floor practise when exercising without the machine.

Preparation 🔍

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist. If no problem, proceed to week 1.

- If you're worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist.

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P03

P03 features long recovery rests between muscle contractions

Phase 1: 15 minutes low frequency.

Phase 2: 30 minutes higher frequency.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Use only alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P03

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or experiencing any type of irritation. If your internal tissues are tender or easily irritated, continue an alternate day pattern.

1 2 3 4 Week 3 - 4: Use P04

Features longer muscle contraction times (8 with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation. Don't worry if you can only join in for a few, let the machine do the rest of the work.



Check in

If Urgency, frequency or urge incontinence are persisting at week 4 add a neuromodulation programme for alternating sessions using P09.

If you're happy using the internal probe, P09 can be applied directly to the pelvic area. OR an alternative method of application using external skin electrodes at the ankle (TTNS) can be used, steps on how to do this can be found on page 47.

Second Month 1 2 3 4

Goal: Improve your muscle (& brain!) skills

1 2 3 4 Week 1 - 4: Use P10

It's time to work yourself a little harder in phase 2 by joining in with the machine more and having shorter rests between contractions.

Phase 1: First enjoy 10 minutes of continuous low-frequency current improves awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes of medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the full 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive)

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases.

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!



Check in

If urgency, frequency or urge incontinence persist continue using P09 for alternating sessions. Steps for using P09 can be found on page 47.

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone?

Do you need to focus another month on strength exercises or more time with the programme that calms the bladder?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the Maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Remember to join in with some of the contractions to keep your brain well connected to the pelvic floor muscle work!

TTNS Alternative

Treatment using only external ankle pads

Transcutaneous Tibial Nerve Stimulation is the specific application of the pre-set neuromodulation programme (P09) using two electrode pads placed at the ankle. The posterior tibial nerve shares a common pathway with the bladder muscle and therefore is an effective treatment for calming a sensitive or overactive bladder.

P09 is a short 20-minute nerve- calming neuromodulation programme for sensitive nerves. A low frequency current has a soothing effect on the bladder nerves.

You will feel the current in cycles of 10 seconds with 5 second rests. Use daily, progressing to twice a day after the first 6 weeks if comfortable, allowing at least 8 hours between treatment sessions. It is recommended to use TTNS for 12 weeks to see improvement in symptoms.

Steps

Empty your bladder before you start using the equipment.

Attach a lead wire to a pair of electrode pads and insert the other end of the lead wire into Channel A on the top left of your device.

Place on the skin - Position one behind the bone on the inside ankle and the other above this at least 3 fingers width apart.



Use Channel A controls to change the intensity of the current through the pads.

- Starting from 0, increase the mA until you feel a slight tingling sensation, you should not feel any discomfort. If you feel discomfort or your big toe twitch, the mA is too high and should be lowered.
- You do not need to feel the muscles of the foot, leg, pelvis or pelvic floor contracting for this programme to be effective.
- Allow the sensations to come and go, relax throughout the programme.
- You can change which ankle you attach the electrodes to each treatment session, especially if you have sensitive or fragile skin.

NOTE: Please do not use this electrode placement or instructions for any other programme.



3. Uncontrollable Gas/Wind and Bowel Control

The anal sphincter muscle is positioned right in the middle of the collective muscles that make up the pelvic floor. If you have symptoms of uncontrollable gas, wind, bowel urgency or frequency (needing to go more than you used to), and faecal incontinence (having an accident before you reach the toilet) you will benefit from exercises that strengthen your pelvic floor muscles.

A STIM machine like the Kegel8® Ultra 20 can be used to help you gain more control. To gain greater strength, you can stimulate the pelvic floor muscles using an electrode inside the vagina (this reaches very close to the anal area as well as all the other muscles) or you can use an electrode designed to go inside the anal canal if you want to target the anal sphincter itself more intensely.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If you're worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist.

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: "Find & Feel" your pelvic floor muscles

1 2 3 4 Week 1-2: Use P03

P03 features long recovery rests between muscle contractions. You can use a vaginal or anal electrode to run this programme. This programme is a gentle way to get started with using stimulation to exercise your muscles.

Phase 1: 15 minutes low frequency.

Phase 2: 30 minutes higher frequency.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Use only alternate days as you get used to the probe.
- If week 1 is comfortable, use your stimulation device daily for 6 days, then have a rest day.
- Take an extra "rest" day if at all uncomfortable or experiencing any type of irritation. or continue with an alternate day pattern.

OPTION: short workout

The muscles of the anal sphincter can feel sensitive

or fatigue easily. This can especially be the case if you are choosing to use an anal probe. You may prefer to use P16 initially which is a short 10 minute exercise programme with stimulation in a 5 second work/ 5 second rest pattern.

1 2 3 4 Week 3 - 4: Use P04

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation. Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.
- Then focus your attention more on the back area of the pelvic floor (as if stopping wind). Again, don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Second Month 1 2 3 4

Goal: Improve your muscle (and brain!) skills

1 2 3 4 Week 1 - 4: Use P17

40 minutes of medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the full 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive)



Check in

If Urgency, frequency or urge incontinence are persisting at week 4 add a neuromodulation programme for alternating sessions using P09.

If you're happy using the internal probe, P09 can be applied directly to the pelvic area. OR an alternative method of application using external skin electrodes at the ankle (TTNS) can be used, steps on how to do this can be found on page 47.

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Continue to use P17

Try some more advanced skills:

- Regularly change your focus from the back part of the pelvic floor and anal sphincter and then using the entire pelvic floor (front, back and sides).
- Aim to hold your muscles with your mind as well as with the machine for the entire contraction part of the cycle. Make sure you are not fading before the contraction is over.
- Make sure you breathe when you contract your muscles.
- Relax properly in the rest part of the pattern.
- Remember to exercise in different positions lying, sitting and now standing.

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. What's still bothering you?

Do you need to focus another month on strength exercises or more time with a programme that calms the bowel nerves?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the Maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Remember to join in with some of the contractions to keep your brain well connected to the pelvic floor muscle work!



4. Menopause, Perimenopause and Postmenopause

Menopause is a stage in a woman's life where she ceases to menstruate. Perimenopause is a 5-10 year phase prior to menopause, where women experience the gradual decline of oestrogen. As you move through these stages of life, you may notice symptoms that suggest your pelvic floor muscles need some attention. This could be bladder or bowel control issues, feeling “unsupported” in the vagina, or a change in your enjoyment of sex.

A STIM device can be used to improve your pelvic floor tissues' health and strength so that you are less inclined to experience problems associated with a weakened pelvic floor. Common issues that can

accompany menopause and perimenopause are:

Stress Incontinence - the leakage women experience when the bladder is under pressure or high impact - like when you cough, sneeze, laugh or jump. Exercising to strengthen and tone the pelvic floor muscles helps improve closure of the bladder tube to prevent leakage.

Pelvic Organ Prolapse - the movement of internal organs downwards in the pelvis. Exercising to strengthen & tone the pelvic floor muscles helps improve the support shelf for your organs and increases your core muscle strength to help with lifting and carrying activities.

Urgency and Frequency of Urination – when you don't feel you can get to the toilet in time or when you feel the need to use the toilet more frequently. Exercising the pelvic floor muscles will improve bladder control.

Sexual Sensation - exercising to strengthen and tone the pelvic floor helps improve the movement you can create and feel during sex. You can also use a STIM device to help you learn to relax as well as to increase blood flow to the vaginal tissues for increased sensation.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P03

P03 features long recovery rests between muscle contractions.

Phase 1: 15 minutes low frequency

Phase 2: 30 minutes higher frequency

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Only use alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P03

If week 1 was comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or experiencing any type of irritation.
- If your internal tissues are tender or easily irritated, continue an alternate day pattern.

1 2 3 4 Week 3 - 4: Use P04

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation.
- Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Second Month 1 2 3 4

Goal: Improve your muscle (and brain!) skills

1 2 3 4 Week 1 - 4: Use P10

It's time to work yourself a little harder in phase 2 by joining in with the machine more and having shorter rests between contractions.

Phase 1: First enjoy 10 minutes of continuous low-

frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity.

Phase 2: 30 minutes of medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive).



Check in

If Urgency, frequency or urge incontinence are persisting at week 4 add a neuromodulation programme for alternating sessions using P09.

If you're happy using the internal probe, P09 can be applied directly to the pelvic area. OR an alternative method of application using external skin electrodes at the ankle (TTNS) can be used, steps on how to do this can be found on page 47.

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases.

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

- Aim to hold your muscles with your mind as well as with the machine for the entire contraction part of the cycle. Make sure you are not fading before the contraction is over.
- Remember to exercise in different positions (lying, sitting and standing). Especially standing now as this is where life takes place!

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the Maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

- Are you working on good bladder habits and doing your pelvic floor exercises without the machine too?
- Remember to join in with some of the contractions to keep your brain well connected to the pelvic floor muscle work!




5. Pelvic Organ Prolapse


Pelvic Organ Prolapse causes discomfort in the vagina or pelvis and can affect bladder control or bladder or bowel emptying.

Doing regular pelvic floor exercise sessions several times a day will build up your muscle tone to make a thicker support shelf for the organs to rest on. The muscles need to be good at holding without tiring and need to be able to react quickly to give extra support when you are lifting and carrying things.

Improving your pelvic floor muscle strength and endurance also helps bladder and bowel control. Learning to release the muscles completely also helps with bladder and bowel emptying.



Well supported internal organ



Mild front wall & back wall prolapse

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P03

P03 features long recovery rests between muscle contractions

Phase 1: 15 minutes low frequency.

Phase 2: 30 minutes higher frequency.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Only use alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P03

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or experiencing any type of irritation.
- If your internal tissues are tender or easily

irritated, continue an alternate day pattern.

① ② ③ ④ Week 3 - 4: Use P04

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation.
- Then focus your attention more on the back area of the pelvic floor (as if stopping wind).
- Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Second Month 1 2 3 4

Goal: Improve your muscle (& brain!) skills

1 2 3 4 Week 1 - 4: Use P10

It's time to work yourself a little harder in phase 2 by joining in with the machine more and having shorter rests between contractions.

Phase 1: First enjoy 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes of medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the full 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive)

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases.

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

TIPS:

- Aim to hold your muscles with your mind as well as with the machine for the entire contraction part of the cycle. Make sure you are not fading before the contraction is over.
- Be extra aware of completely relaxing in the rest part of the cycle.
- Are you breathing all through the cycle?
- During the contractions - could you move your arms (try some biceps curls, or backstroke) and hold your muscles at the same time and breathe?!

- Change the positions in which you practise. You could do one session lying, one session sitting, and if you have no problem retaining the probe, one session standing (but stand STILL - don't try walking about!)

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone? What's still bothering you?

Are you working on good bladder & bowel habits, paying attention to your lifting and carrying techniques and doing your pelvic floor exercises without the machine too?

Not making the progress you hoped? Reach one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the maintenance programme!

Maintenance ↻

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Phase 1: 15 minutes.

Phase 2: 30 minutes.

- Are you working on good bladder habits and doing your pelvic floor exercises without the machine too?
- Remember to join in with some of the contractions to keep your brain well connected to the pelvic floor muscle work!



6. Pelvic Pain

If you're suffering from pelvic pain or anal or vaginal discomfort (including tightness and endometriosis), using a STIM device like the Kegel8® Ultra 20 can bring you relief.

Vagina walls are layered with the pelvic floor muscles at the base and the sides. These muscles need to have good tone and firmness to make the space feel supported. They need to be able to contract strongly for bladder and bowel control when needed.

However, the muscles only need to work just “enough” in everyday life to be part of our supportive posture. They do not need to “grip”, or we can get a cramp or spasm.

Like with shoulders, backs, and hamstrings, there must be a balance between useful muscle work and “overwork”. Muscles that work too much without being rested and stretched can become short or “tight”. We now also understand that the pelvic floor muscles must be able to “switch off”, stretch out and relax fully, both spontaneously and on command. Muscles that can’t switch off tire easily and, as a result, offer poor bladder control.

Muscles that are overworked can also become painful as well as make the pelvic organs feel compressed. This pain can be localized to just the vagina or sometimes felt in the surrounding abdomen and other parts of the pelvis. Pelvic floor muscles need to be able to properly relax to be able to fully empty the bladder and bowel and for sex to be comfortable.



A STIM machine can be used with different settings to achieve different goals. You can use a programme to practise awareness of both contracting and relaxing the pelvic floor muscles. Stimulation can also be used to activate helpful chemical reactions in the brain, which help to reduce pain.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

Relaxation

Goal: To relax your pelvic floor muscles and bring blood flow to the area.

P20: Use with vaginal probe

20 minutes of low-frequency current, in two 10-minute phases, improves awareness of the muscle position, blood flow to the tissues, and nerve sensitivity. Lie down, and relax as the low-frequency stimulation brings blood flow to the area.

Phase 1: Sensations will come and go in short 3 second cycles with 7 second rest periods. Take time to notice the sensations sent to your brain from this area. You may find that you get “used” to the sensations after 5-10 minutes and can turn the mA up a little higher. Aim to feel tingling/knocking sensation, once you feel this you are at a good level.

Phase 2: Sensations will be continuous. Start to notice all of your body, try to relax your shoulders, legs and back muscles too. 20 minutes is a good time to boost blood flow while you switch “off” your body tension. It will be difficult at first! You might consider combining this pelvic floor programme with mediation or mindfulness (before or after) if relaxing is something you need to “learn” to do.

P19: Use with vaginal probe or surface electrodes

60 minutes of continuous low-frequency current if you want to relax for longer without a change of phase.

Pain Relief

Goal: for pain relief

P18: Use with surface electrodes

This is the same setting as you'd find on a transcutaneous electrical nerve stimulation (TENS) machine, a device similar to STIM, which is frequently used as an alternative to medication. This helps relieve the type of pain often associated with periods, vulvodynia, and general pain in the pelvic area. The high frequency current diminishes your awareness of sensations coming from other places in your body. This stimulation programme is applied to the skin surface using the electrode pads on your lower back/sacrum or abdomen.

P19: Use with vaginal probe or surface electrodes

This programme has a similar effect and purpose but uses a different physiological effect to P18,

(termed acupuncture-like TENS).

The low frequency electrical current stimulates the release of pain-relieving chemicals from your brain. It can be applied with skin electrodes or the internal probe if this feels right for you. A continuous low-frequency current improves awareness of the muscle position, blood flow, and nerve sensitivity. Relax completely for up to 60 minutes.

P20: Use with the vaginal probe

This programme has 2 phases. It can help to relieve the pain caused by overworking your muscles, especially if you feel you want to turn them off. This type of pain often feels like a burning or sharp sensation.

Phase 1: 10 minutes of work/rest pattern to help you learn better when you are contracting your muscles and to notice when they are relaxing. Be mindful of the sensations and try to work with the machine to practise both skills.

Phase 2: 10 minutes of continuous low-frequency current improves awareness of the muscle position, blood flow, and nerve sensitivity. Relax completely.

Awareness

Goal: Exercise to find and feel the pelvic floor muscles.

1 2 3 4 Week 1: Use P06

This 20-minute muscle exercise programme features long recovery rests between muscle contractions.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Use only alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P06

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or irritable internally.
- If your internal tissues are tender or easily irritated, continue an alternate day pattern.

1 2 3 4 Week 3 - 4: Use P10

Phase 1: 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity.

Phase 2: 30 minutes of medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

Help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles. Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation. Don’t worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

- Try to relax completely for the full 5 seconds between the contractions.
- Take breaks whenever you feel tired and allow the machine to do all the work for a few minutes while you allow yourself to feel the sensations.
- If these contractions feel too “quick”, stay a week or so longer with P06.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive).

Second Month 1 2 3 4

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases. Still paying lots of attention to relaxing fully between each contraction.

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

Then use P19

Finish your sessions with 10-20 minutes relaxing with the circulation programme.

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone? What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Then use: P20

Choose to relax and refocus on sensations for 20 minutes once a week, and anytime you are feeling tense, stressed or disconnected.



Precaution:

Less than 12 weeks since giving birth:

Do pelvic exercises without using a STIM device.

More than 12 weeks since giving birth:

Safe to follow this programme.

7. Postpartum Recovery

After giving birth, your pelvic floor muscles will be weakened and stretched by pregnancy and delivery. Doing regular pelvic floor exercise sessions several times a day will build up muscle strength and endurance (which you need for good bladder and bowel control). It will also provide you with the muscle tone you need to prevent prolapse, support your organs, and improve sexual sensation.

In the first 12 weeks after your baby is born, you're encouraged to do pelvic floor exercises but, as the body is still very fragile, it's not advised to use a muscle stimulation machine until after 12 weeks have passed. Until then, do traditional kegel exercises without the assistance of a device. After 12 weeks, you can use a muscle stimulation machine to help improve circulation to the pelvic floor and perineum and to help exercise the muscles.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month ①②③④

Goal: “Find and Feel” your pelvic floor muscles

①②③④ Week 1: Use P14

A 25-minute gentle start exercise programme for vulnerable mums after childbirth.

Phase 1: 20 minutes of continuous low-frequency current improves awareness of the muscle position, blood flow and nerve sensitivity. Relax in this phase.

Phase 2: 15 minutes higher frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 6 seconds and then relaxing for 18 seconds. Long rest phases are needed when your muscles are learning to work again.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Use only alternate days as you get used to the probe.

1 2 3 4 Week 2: P14

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

Take an extra “rest” day if at all uncomfortable or irritable internally.

If your internal tissues are tender or easily irritated, continue an alternate day pattern.

1 2 3 4 Week 3 + 4 Use P03

P03 exercises your muscles for a longer 45 minutes, with a shorter recovery time between muscle contractions.

Phase 1: 15 minutes low frequency.

Phase 2: 30 minutes higher frequency.

Second Month 1 2 3 4

1 2 3 4 Week 1 - 4: Use P04

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation.
- Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P10

It's time to work yourself a little harder in phase 2 by joining in with the machine more and shorter rests between contractions.

Phase 1: First enjoy 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the entire 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive).

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone? What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus.

Doing well - fantastic - move on to the maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Then use: P20

Choose to relax and refocus on sensations for 20 minutes once a week, and anytime you are feeling tense, stressed or disconnected.



8. Vaginal Looseness and Lack of Sensation

Vaginal looseness and lack of vaginal sensation are usually closely linked. Women don't have that many nerve endings inside the vagina. Women feel "pressure" and "deep touch" more than light sensations. The sexual pleasure women experience comes from the rubbing and build-up of friction against the vaginal walls.

The vaginal walls at the base and the sides are layered with the pelvic floor muscles. These muscles need to have good tone and firmness to

make the space feel tight and close to your partner, for the pleasure nerve endings to get activated and for you to be able to grip your partner (for more pleasure for both of you).



A STIM device can be used to both exercise the pelvic floor muscles (to help them grow thicker and stronger) and bring more blood flow to the area to improve nerve sensitivity and general awareness.

In this routine, we alternate between settings, first using a programme to make the muscle fibres repeatedly contract and release to encourage the muscle fibres to grow. The machine targets both the quick and slow muscle fibres, just like you need to do when you practise kegel exercises without a machine. Join in with some of these contractions to teach your brain how to work the muscles when you are not connected.

On another setting there is no contraction of the muscles, but the stimulation triggers nerve activity and increased blood flow - this is a good time to relax and “tune in” to the sensations that you can receive from the vagina area and become more body aware again.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P06

This 20-minute muscle exercise programme features long recovery rests between muscle contractions.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Use only alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P06

If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.

- Take an extra “rest” day if at all uncomfortable or irritable internally.
- If your internal tissues are tender or easily irritated, continue an alternate day pattern.

1 2 3 4 Week 3 - 4: Alternate days with P06 & P19

Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.

- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation.
- Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

P19

Relax as the low-frequency stimulation brings blood flow to the area.

- Take time to notice the sensations sent to your brain from this area. You may find that you “get used” to the sensations after 5-10 minutes and can turn the mA up a little higher.
- Aim for a strong but comfortable sensation. This programme lasts 60 minutes, but just 20 minutes is adequate to boost blood flow.

Second Month 1 2 3 4

1 2 3 4 Week 1 - 4: Use P10

This is a slightly shorter programme, but it's time to work yourself a little harder in phase 2 by joining in with the machine more and shorter rests between contractions to encourage your muscles to keep growing.

Phase 1: First enjoy 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity. Relax in this phase.

Phase 2: 30 minutes medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the full 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive).

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases.

Phase 1: Focus on the back area of the pelvic floor (as if stopping wind). Practise this for 10 minutes.

Phase 2: 30 minutes - When you join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

P19

Finish your sessions with 10-20 minutes relaxing with the circulation programme.

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone? What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Phase 1: 15 minutes.

Phase 2: 30 minutes.

Feeling stressed or disconnected?

Choose P19 to relax and refocus on sensations.



Precaution:

Less than 12 weeks since you had surgery:

Do pelvic exercises without using a STIM device.

More than 12 weeks since you had surgery:

Safe to follow this programme.

9. Rehabilitation After Gynaecological Surgery

Many women need surgery because of a bladder, bowel or prolapse issue. In some cases, a woman may have been working on her pelvic floor muscle exercises before an operation.

After surgery you will want to get back to your normal routine as soon as possible. However, gynaecological surgery can have an incredible impact on pelvic floor muscles, with women often finding that they feel weak, unsupported or de-conditioned in the weeks following their operation.

This is partly because the surgical work in the area affects the blood flow, nerves, and surrounding tissues; you will also be a bit sore and less active for a few weeks during recovery.

Doing regular pelvic floor exercise sessions several times a day, straight away after your surgery, will help the healing process by improving blood circulation to the area, reducing swelling, and strengthening your muscles so they can support you.

The next step is to build up good muscle strength and endurance again which you need for secure bladder and bowel control. Exercises will also improve the muscle tone which we need to prevent pelvic organ prolapse, to feel supported.

In the first 12 weeks after your surgery you should do pelvic floor exercises, but as the body is healing and sensations can be altered, it is not advised to use a muscle STIM device. After 12 weeks post-op, the Kegel8® Ultra 20 can be used to exercise the pelvic floor muscles.

Preparation

IMPORTANT: Before proceeding with this condition exercise, make certain to run built-in programmes P01 and P02. These are sensitivity tests to help you determine if it's safe for you to move forward with the following exercise. (see page 19 - 20).

Prog. No.	Screen Codes	Programme Name	Time (mins)	Page
P01	RCV1	Sensation Test 1	30	19
P02	RCV2	Sensation Test 2	60	20

If during the test you notice a problem with sensation, do not proceed. Instead, arrange a visit with your doctor or physiotherapist.

- If no problem, proceed to week 1.
- If worried that there is a problem with sensation, arrange an appointment with your GP or a specialist physiotherapist

TIP: For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

First Month 1 2 3 4

Goal: “Find and Feel” your pelvic floor muscles

1 2 3 4 Week 1: Use P03

Features long recovery rests between muscle contractions.

Phase 1: 15 minutes of low frequency.

Phase 2: 30 minutes higher frequency.

- Allow the sensations to come and go in cycles of stimulation and rest periods. Aim to learn where your pelvic floor muscles are.
- Get used to finding a comfortable mA setting.
- Use only alternate days as you get used to the probe.

1 2 3 4 Week 2: Use P03

- If week 1 has been comfortable, use your stimulation device daily for 6 days, then have a rest day.
- Take an extra “rest” day if at all uncomfortable or irritable internally.
- If your internal tissues are tender or easily

irritated, continue an alternate day pattern.

1 2 3 4 **Week 3 + 4 Use P04**

Features longer muscle contraction times (8 seconds) with a similar long rest time in each cycle. Four phases to change between stimulation frequencies more often.

- Now help your brain to copy the machine work by intermittently “joining in” with some of the stimulation cycles.
- Try to work your pelvic floor muscles with your mind, drawing in and up when you feel the stimulation.
- Don't worry if you can only join in for the first few seconds. Let the machine do the rest of the work.

Second Month ①②③④

Goal: Improve your muscle (and brain!) skills

①②③④ **Week 1 - 4: Use P10**

It's time to work yourself a little harder in phase 2 by joining in with the machine more and shorter rests between contractions to encourage your muscles to keep growing.

Phase 1: First enjoy 10 minutes of continuous low-frequency current to improve awareness of the muscle position, blood flow, nerve sensitivity.

Phase 2: 30 minutes medium frequency current to stimulate the fast and slow pelvic floor muscle fibres by contracting for 5 seconds then relaxing for 5 seconds in cycles.

- Join in with some of these contractions, trying to contract for the full 5 seconds.
- Join in for 1 minute, then let the machine do the work for 1 minute or count joining in for 6 cycles, then letting the machine do 6 cycles.
- Use 6 days out of 7 (or alternate days if your internal tissues are sensitive)

Third month 1 2 3 4

Goal: Develop more pelvic floor skills

1 2 3 4 Week 1 - 4: Use P11

This month join in with some of the contractions in both phases.

Phase 1: 10 minutes - focus on the back area of the pelvic floor (as if stopping wind).

Phase 2: 30 minutes - when join in try to use ALL of the muscle, the back area, the front around the bladder tube and even the side walls!

Symptom review

Maintenance or seek more help?

How are you getting on? Compare how you feel now to how you felt in Week 1. Are some problems and worries gone? What's still bothering you?

Not making the progress you hoped? Reach out to one of our advisors by visiting our website:

www.kegel8.co.uk/contactus

Doing well - fantastic - move on to the maintenance programme!

Maintenance

Maintain your results!

Use P05: Weekly maintenance.

A mix of frequencies to maintain fast and slow twitch muscle fibres.

Then use: P20

For pelvic tightness and vaginal tightness, use the vaginal probe with P20 on a very low frequency to help relax your muscles.

Frequently Asked Questions

We hope that this guide has helped you get familiar with the Kegel8® Ultra 20. Please refer to the Technical Instructions for Use (IFUs) that accompanied this unit for further details about exercise programmes and caring for your device.

Screen & Display questions

On the LCD screen what does mA stand for?

MA is an abbreviation for milliamperere or milliamp. A milliamp is a unit of measurement equal to one-thousandth of an amp. An amp or ampere is a unit of electrical current. By using such a small current you are able to safely operate the Kegel8® Ultra 20 from the comfort of your home while still attaining significant results.

Why are there two channels on the top of the device?

There are two channels because the Kegel8® Ultra 20 enables you to use different ways to apply the current. Channel A (left- hand side) is for a vaginal probe and Channel B (right-hand side) is for an anal probe or attaching electrodes. The two channels also allow you to use up to 4 electrode pads at the same time. We have a range of electrode pads, vaginal probes, and anal probes available online at www.kegel8.co.uk.

Why are there two mA numbers at the bottom of the LCD screen?

The two large numbers at the bottom of the display represent the amount of electricity being output through Channel A and Channel B of your device.

Programme questions

How do I know which programme to use?

Your Kegel8® Ultra 20 comes complete with a library of built-in exercise programmes relieving you of the need to guess which programme to use. Our table on page 17 - 18 will help you choose or browse our section with advice from a specialist physiotherapist about how to use your machine best to help with 9 of the common medical conditions.

How do I know how high to turn the mA setting?

The amount of current needed to stimulate pelvic floor muscles varies from person to person. Starting from 0mAs, you should increase the mA power to a level where you can feel the effects of the electrical pulse without it causing any discomfort.

For the first week of your routine, increase the mA until you feel a slight tingling sensation. After this, increase your mA's 1-2 levels every 1-2 weeks. If it becomes uncomfortable, lower it again.

When conducting my pelvic muscle exercises I

aim to use my Kegel8® Ultra 20 at mA 40 but on some programmes it drops to 30 - why is that?

While the programme goes through its preset routine, the rate and length of the electrical pulses vary. Some pre-set programmes have 2 or more phases. The rate (Hz) and length (μ S) of the electrical impulses will vary. When the phase changes, the mAs drop (or stop) so that the new type of current does not take you by surprise. If you feel a decrease in intensity, or no current at all, just use the + button to increase the mAs back up to a comfortable level.

Why can't I get the mAs to increase past 6, 7 or 8?

This is an issue that is usually caused by the probe not making effective contact with your muscles. We recommend that you try using your Kegel8® Ultra 20 while lying down or sitting with your legs elevated. This will help to prevent the probe from falling out of position. If the probe's metal strips are not in direct contact with your muscles, the mAs cannot go higher and the current cuts off for your safety. You can also get better results by using a water-based lubricant, such as KE Gel.

This could also be an issue with your probe or lead wire. To test your probe and lead wire, please visit www.kegel8.co.uk/help/testing-your-kegel8 for further instructions.

Probe & Battery questions

Why does the probe twist or not stay in place?

To work correctly, the metal strips on your probe need to be in direct contact with your muscles. We recommend that you sit with your legs held tightly together, or to use your Kegel8® Ultra 20 while lying down or with your legs elevated.

Why can I only feel muscle stimulation on one side?

This could be due to:

The position of the probe. To work correctly, the metal strips on your probe need to be in direct contact with your muscles. We recommend that you sit or lie with your legs close together, to bring the pelvic floor muscles as close as possible against the probe. You can also get better results by using a water-based lubricant such as KE Gel.

You may have nerve damage to one side. This type of nerve injury can occur during childbirth (caused by the pressure from the baby's head during delivery). It means you have less sensation on one side. If this is the case, we recommend working off of the side you can feel; do not increase the mA's higher to try and feel both sides. Instead, keep it at a level which is comfortable for the side with sensation, the device will help to bring back sensation in the other side over time.

How far should I insert the vaginal probe?

Avoid placing the probe too deeply, the front end should be inserted no further than 2 inches (5.5 cm) into your vagina. It should not press on your cervix.

I find it painful inserting the probe. What should I do?

Try using a water-based lubricant such as KE Gel.

If this does not help, see your doctor as you may be suffering from vaginal atrophy, which causes dryness and thinning of the vaginal walls. It can be treated quite easily, so don't worry, you will soon be able to use your Kegel8® Ultra 20.

How long will the 9v battery that came with my device last?

If you use the Kegel8® Ultra 20 daily, the battery should last about 3 months. You can get replacements from us or at your local supermarket - please choose a quality 9v battery that will not leak. You will know when your battery is low because a battery icon will flash on your Kegel8® Ultra 20 screen.

How long will my probe last?

With normal use, your probe should last between 12 - 18 months. We recommend you replace your probe after this time to ensure you continue to get the best possible results from your Kegel8® Ultra 20.

When to use your Kegel8 Ultra 20

How soon after the birth of my baby can I start with my Kegel8® Ultra 20?

In the first 12 weeks after your baby is born, you're encouraged to do pelvic floor exercises to help you regain strength but, as the body is still very fragile, it's not advised to use a muscle stimulation device like the Kegel8® Ultra 20 until after 12 weeks have passed.

If I have an IUD, Mirena coil, or pessary, can I use this device?

Yes. The Kegel8 Ultra 20 does not affect any of these devices.

Can I use Kegel8 Ultra 20 while I am pregnant?

No. Clinical guidelines advise against using any vaginal STIM devices during pregnancy.

I have had pelvic surgery. How soon can I start using the Kegel8 Ultra 20?

Using the Kegel8 Ultra 20 after pelvic surgery can help to speed recovery, however, you need to wait at least 12 weeks after surgery before using the device. Every case is unique so make certain to first consult with your doctor.


Can I use my Kegel8 Ultra 20 when on my period?

Yes, however, if blood flow is heavy, you may prefer to avoid using the vaginal probe until your flow is lighter.

Product Support

We hope that this guide has helped you get familiar with the Kegel8® Ultra 20. Please refer to the Technical Instructions for Use (IFUS) that accompanied this unit for further details about exercise programmes and caring for your device.

If at any moment you have a question and cannot find the answer in this guide, we are happy to assist you.



You can reach one of our specialist Kegel8 advisors by booking a telephone appointment or live chat by visiting our website www.kegel8.co.uk/schedule-a-call We hope you enjoy using your Kegel8® Ultra 20!

More than just a retailer

With extensive experience in women's health, we have helped thousands of women to overcome their problems with clinically backed solutions.